Print on plain white paper - Part # 122919 - (CUT INTO HALF)

INSTRUCTIONS for kit # 152816

Secondary Power Hook-up for Universal Light

WARNING - Disconnect all equipment from the power supply before attempting this installation. Failure to do so represents a potential shock hazard with possible personal injury.

1. Run the light cable along the bottom of the chair mount
2. Secure the light cable to the bottom of the chair mount using three plastic cable clamps provided.
3. Feed the light cable into the back of the chair.
4. Secure the light cable to exiting wiring on the lower lift arm with the cable tie provided.
5. Remove the side cover.
6. Feed the light cable through the hole in the upright housing.
7. Make the black and white wire connections as shown.
8. Add the ground jumper to the upright housing and connect it to the light cable ground wire. Use the wire nut provided.
9. Replace all covers and hardware.







SOME DETAILS REMOVED FOR CLARITY

Knight®
by Midmark

Midmark Corporation • P O Box 286 • 60 Vista Drive
Versailles • Ohio • 45380-0286 • www.midmark.com
Phone: 937-526-3662 Fax: 937-526-5542

TPW1520 Rev. B

INSTRUCTIONS for kit # 152816

Secondary Power Hook-up for Universal Light

WARNING - Disconnect all equipment from the power supply before attempting this installation. Failure to do so represents a potential shock hazard with possible personal injury.

1. Run the light cable along the bottom of the chair mount
2. Secure the light cable to the bottom of the chair mount using three plastic cable clamps provided.
3. Feed the light cable into the back of the chair.
4. Secure the light cable to exiting wiring on the lower lift arm with the cable tie provided.
5. Remove the side cover.
6. Feed the light cable through the hole in the upright housing.
7. Make the black and white wire connections as shown.
8. Add the ground jumper to the upright housing and connect it to the light cable ground wire. Use the wire nut provided.
9. Replace all covers and hardware.







SOME DETAILS REMOVED FOR CLARITY

Knight®
by Midmark

Midmark Corporation • P O Box 286 • 60 Vista Drive
Versailles • Ohio • 45380-0286 • www.midmark.com
Phone: 937-526-3662 Fax: 937-526-5542

TPW1520 Rev. B